



中国合格  
国际互认  
检测  
TESTING  
CNAS L16167

Test Report

No: GZCPCH200401472

Date: 2020-04-20

Client name:  
Client address:



Sample name:  
Batch No./Date:  
Manufacturer:

Doctor care disposable protection mask  
AL021/2021/04/02  
Foshan Cherish Sanitary Goods Co Ltd

Above sample(s) was/were submitted and certified by the client, SGS quoted the information with no responsibility as to the accuracy, adequacy and/or completeness.

SGS job No.: GZCPCH200401472  
Date of receipt: 2020-04-08  
Testing period: 2020-04-08--2020-04-20

**TEST(S) REQUESTED:**

Selected test(s) as requested by applicant:  
Please refer to next page(s).

**TEST METHOD(S):**

Please refer to next page(s).

**TEST RESULT(S):**

Please refer to next page(s).

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, and this document cannot be used for publicity without approval of the Company. It's not be allowed to copy testing report (except for copy of full text) without written approval.

Signed for and on behalf of  
SGS CSTC Standards Technical Services Co., Ltd. Guangzhou Branch  
Approved Signatory



Approved Signatory

SGS-CSTC Standards Technical Services Co., Ltd. Guangzhou Branch

Page 1 of 2

RAND: 2594316



SGS-CSTC Standards Technical Services Co., Ltd. Guangzhou Branch  
Address: 11th Floor, Zhongshan Road, Guangzhou, China  
Tel: 86-20-37551111 Fax: 86-20-37551112  
E-mail: gzs@sgs.com.cn  
Website: www.sgs.com.cn  
Member of the ISO 9001:2015 and ISO 14001:2015 certified organizations



由 扫描全能王 扫描创建

Test Report

No: GZCPCH200401472

Date: 2020-04-20

**TEST METHOD(S):**  
GB 15979-2002 Hygienic standard for disposable sanitary products

**TEST RESULT(S):**

Test item(s)	Unit	Test method(s) (Reference to)	Limit <sup>s</sup>	Test result(s)	Single Judgment
Aerobic Plate Count	CFU/g		≤ 200	< 20	PASS
Fungi	CFU/g		≤ 100	< 20	PASS
Coliforms	/0.25g	GB 15979-2002	Absence	Not detected	PASS
Staphylococcus aureus	/0.25g		Absence	Not detected	PASS
Streptococcus Hemolyticus	/0.25g		Absence	Not detected	PASS
Pseudomonas aeruginosa	/0.25g		Absence	Not detected	PASS

Remark: \* = Limit from GB 15979-2002 Hygienic standard for disposable sanitary products.

**SAMPLE DESCRIPTION:**

Sample in bag

Photo Appendix



\*\*\* End of Report\*\*\*



SGS-CSTC Standards Technical Services Co., Ltd. Guangzhou Branch  
Address: No. 100, Zhongshan Road, Guangzhou, Guangdong, P.R. China  
Tel: +86-20-37331842 Fax: +86-20-37331843  
E-mail: gzs@sgs.com.cn  
Website: www.sgs.com.cn  
SGS-CSTC Standards Technical Services Co., Ltd. Guangzhou Branch  
Address: No. 100, Zhongshan Road, Guangzhou, Guangdong, P.R. China  
Tel: +86-20-37331842 Fax: +86-20-37331843  
E-mail: gzs@sgs.com.cn  
Website: www.sgs.com.cn

